

HYDRAULIC IMPULSE CALIBRATOR

PERFORMANCE (ACTUATOR)

Manufacturer Model Number	PCB	913B02	[1]
Amplitude Range (useable)	psi [Bar]	200 to 15,000 [13.8 to 1034.2]	
Typical Rise Time	Ms	3	[2]
Typical Pulse Duration	Ms	6 to 8	[2]

PERFORMANCE (REF TRANSDUCER)

Manufacturer Model Number	PCB	136A	[1]
Sensor Type		Charge Piezoelectric	
Sensing Element Material		Tourmaline	
Sensitivity (+/- 15%)	pC/psi [pC/kPa]	0.2 [0.029]	
Measurement Range	psi [Bar]	15,000 [1034.2]	
Resolution	psi [kPa]	0.5 [3.45]	[3]
Resonant Frequency	MHz	≥ 1.0	
Rise Time (Reflected)	μs	≤ 3.0	
Linearity	% FS	≤ 0.5	
Capacitance	pF	10	

PERFORMANCE (CHARGE AMPLIFIER)

Manufacturer Model Number	PCB	443B02	[1]
Time Constant	s	10 to 100,000	[4][5]
Frequency Response (-10%)			
High Pass	Hz	0.2 or 2.0	
Low Pass	kHz	0.1, 1, 3, 10, 30, 100 or 200	
Input Range	pC	up to 100,000	
Charge Sensitivity	V/pC	0.0001 to 10	
Broadband Noise (1 Hz – 10 kHz)	μV [dB]	9 [-101]	[6]
Drift	pC/s	< 0.03	[7]

PHYSICAL

Two-port Test Manifold	Material	Brass	
Drop Tube	Material	Aluminum	[8]
Base Assembly	Material	Aluminum with Delrin Pad	
Impact Mass	Material	Stainless Steel	[9]
Piston with O-ring	Material	Stainless Steel	
Oil Couplant	Material	Dow Corning 200	[10]
Shop Rags	Material	Cotton	
Oil Couplant	Material	Dow Corning 200	[10]
Shop Rags	Material	Cotton	

All specifications are at room temperature unless otherwise specified.

ICP® is a registered trademark of PCB Piezotronics, Inc.


Microsoft, Access, and Excel are registered trademarks of Microsoft Corporation.

In the interest of constant product improvement, specifications may change without notice.

K9913C CALIBRATION SYSTEM COMPONENTS*

Qty 1	TMS 9913 Hydraulic Impulse Calibrator
Qty 1	PCB 108A04 Verification Pressure Transducer
Qty 1	PC System Controller, includes PC, Flat Panel Monitor, and Printer
Qty 1	National Instruments PCI 4461 24-bit Data Acquisition Card
Qty 1	PCB 482A21 ICP Sensor Signal Conditioner
Qty 1	TMS Automated Impulse Calibration Software
Qty 1	Microsoft Access for Transducer Database
Qty 1	Microsoft Excel for Certificate Print Generation

PCB 913B02 two-port manifold shown assembled in base with piston and transducers installed, in impact-ready position.



* System also available as option to TMS 9155 Accelerometer Calibration Workstation (reference model 9155D-913)

NOTES:

- [1] See manufacturer data for full specifications
- [2] Measured using Dow Corning 200 silicon oil
- [3] Typical value, dependent upon range setting and cable length used in charge amplifier system
- [4] Dependent upon gain setting
- [5] Medium TC mode 10, 100, or 1000 s. Long TC mode up to 100,000 s
- [6] Measured at gain of 1mV/pC with 1nF cap
- [7] In long TC mode only
- [8] 3 ft (91 cm) in length
- [9] Variety of sizes, configurable up to 9 lb (4 kg)
- [10] 120 mL

Accessories:

- 1x 003C10 cable
- 1x 012A03 cable
- 1x 061M149 adapter
- 1x 061M150 adapter
- 1x 061M151 adapter
- 10x 065A06 ¼" seal
- 10x 065A09 ½" seal
- 1x sieve

Project Engineer:

Product Manager:

Mkt Team Leader:

Spec Number:

PS-0058

Date: 8-16-12

Date: 8/20/12

Date: 8/16/12